

Effects of hemolysis interferences on routine biochemistry parameters

Mehmet Koseoglu, Aysel Hur, Aysenur Atay, Serap Çuhadar*

Ataturk Training and Research Hospital, Department of Biochemistry and Clinical Biochemistry, Izmir, Turkey

*Corresponding author: sdcuhadar@yahoo.com

Abstract

Introduction: Hemolysis is still the most common reason for rejecting samples, while reobtaining a new sample is an important problem. The aim of this study was to investigate the effects of hemolysis in different hemolysis levels for mostly used biochemical parameters to prevent unnecessary rejections.

Materials and methods: Sixteen healthy volunteers were enrolled in the study. Four hemolysis levels were constituted according to hemoglobin concentrations and they were divided into five groups: Group I: 0-0.10 g/L, Group II: 0.10-0.50 g/L, Group III: 0.51-1.00 g/L, Group IV: 1.01-2.50 g/L, Group V: 2.51-4.50 g/L. Lysis was achieved by mechanical trauma.

Results: Hemolysis interference affected lactate dehydrogenase (LD) and aspartate aminotransferase (AST) almost at undetectable hemolysis by visual inspection (plasma hemoglobin < 0.5 g/L). Clinically meaningful variations of potassium and total bilirubin were observed in moderately hemolyzed samples (hemoglobin > 1 g/L). Alanine aminotransferase (ALT), cholesterol, gamma glutamyltransferase (GGT), and inorganic phosphate (P) concentrations were not interfered up to severely hemolyzed levels (hemoglobin: 2.5-4.5 g/L). Albumin, alkaline phosphatase (ALP), amylase, chloride, HDL-cholesterol, creatine kinase (CK), glucose, magnesium, total protein, triglycerides, unsaturated iron binding capacity (UIBC) and uric acid differences were statistically significant, but remained within the CLIA limits.

Conclusion: To avoid preanalytical visual inspection for hemolysis detection, improper sample rejection, and/or rerun because of hemolysis, it is recommended in this study that, routine determination of plasma or serum free hemoglobin concentrations is important. For the analytes interfered with hemolysis, new samples have to be requested.

Key words: hemolysis; preanalytical errors; interference; analytes.

Received: October 1, 2010

Accepted: December 28, 2010

Introduction

Hemolysis is the most common preanalytical source of error in clinical laboratories and responsible for nearly 60% of rejected samples (1,2). It may occur both *in vivo* or *in vitro*. The ratio of *in vivo* hemolysis is only 3.2% of all the hemolyzed specimens (3). *In vitro* hemolysis occurs more often and it is caused by improper sample drawing, handling or centrifugation. Especially hardly collected samples, or stored and/or transported, have increased risks for hemolysis.

Most of the hemolyzed samples are being rejected on pre-analysis stage according to the visual

detection of serum interferences, even if the requested tests may not be interfered with hemolysis. Besides, according to the reports, visual assessment of sample hemolysis showed little agreement with the actual concentration of hemoglobin interferent (4-7).

Conversely, even if the hemolysis is not visible, there is also a discharge of the cell constituents into serum or plasma (8). So invisible hemolysis is an important cause of false results and has to be detected before the investigation procedure. Increased number of biochemical tests, number of sam-

ples and increased automation obligated laboratory staff to study harder, give results faster, so pre-analytical determination of hemolyzed sample and determination of interfered analytes before analysis became more important.

Since the knowledge of possible effects is important for correct interpretation of the results, the aim of the present work was to evaluate the effects of hemolysis interferences in different levels on routine commonly used biochemistry tests. Results of this study may help to prevent unnecessary rejections of samples.

Materials and methods

The venous blood samples, collected from 16 healthy volunteers, were taken into 5 different heparinized tubes (Vacuette®, Greiner Labortechnik, Germany) to study the effect of *in vitro* hemolysis. Four samples were drawn through the needles of 5 mL syringe (1.5 inch, 21 gauge) speedily for 2, 4, 6 and 8 times respectively to lyse the cells by mechanical trauma to obtain slightly, mildly, moderately and severely hemolyzed samples. They were all centrifuged at 1000 x g for 15 minutes. This method of cell lysis was chosen because blood transferring into a tube by pushing forcibly down on the syringe plunger is analogous to the mechanical disruption of erythrocytes that frequently occurs during blood collection. In this method there is no standardization way of the force applied while transferring the blood by syringe. Besides every patients' fragility of red blood cell is different, so free Hb concentrations of all samples were not correlated with the force.

After measuring free Hb concentrations of the hemolyzed samples, the samples were grouped according to their free Hb concentrations. The exclusion criteria of a hemolyzed sample was the concentration of its free Hb concentration discordant with the degree of hemolysis determined for each group, so the number of samples were not equal at each group.

Free Hb of the samples were measured spectrophotometrically (Shimadzu Corporation; Kyoto, Japan) with Na₂CO₃ solution (10 mg/100 mL) as a re-

agent (9). Absorbances were measured at 415, 450 and 700 nm for all hemolyzed and diluted plasma. Total plasma hemoglobin was calculated according to the formula:

$$\text{Hb} = 154.7 \times (\text{A425}) - 130.7 \times (\text{A450}) - 123.9 \times (\text{A700}) \quad (9).$$

(Reference ranges: 0-0.1 g/L for plasma free Hb).

Specimens were categorized according to the Hb concentrations into 5 groups:

Group I (0-0.1 g/L), nonhemolyzed (N = 15);

Group II (0.10-0.50 g/L) slightly hemolyzed (N = 12);

Group III (0.51-1.00 g/L) mildly hemolyzed (N = 10);

Group IV (1.01-2.50 g/L) moderately hemolyzed (N = 15);

Group V (2.51-4.5 g/L) severely hemolyzed (N = 12).

The results were analyzed to determine if slight, mild, moderate or severe hemolysis had a significant impact on the analyte concentrations.

All analytes were measured with Olympus analyzer with original reagents (Olympus AU2700 system reagent, Olympus Diagnostica GmbH, Lismeehan, O'Callaghan's Mills, Co. Clare, Ireland).

Plasma concentrations of albumin, alkaline phosphatase (ALP), alanine aminotransferase (ALT), α-Amylase, aspartate aminotransferase (AST), total bilirubin, calcium, sodium, potassium, chloride, total cholesterol (TC), creatine kinase (CK), creatinine, gamma glutamyltransferase (GGT), glucose, high density lipoprotein cholesterol (HDL-C), iron (Fe), lactate dehydrogenase (LD), magnesium, inorganic phosphate, total protein, triglyceride (TG), unsaturated iron binding capacity (UIBC), urea, uric acid were analyzed in all groups.

The effects of the hemolysis were evaluated according to the total allowable error recommendations of Clinical Laboratory Improvement Amendments (CLIA'88) (10) (Table 1). CLIA' 88 regulations have established fixed limits for assessing method and laboratory performance for specific regulated analytes. In practice, the total allowable error for a given analytical method must be less than the res-

pective CLIA fixed limits for the analyte in question.

Statistical analysis

To compare the concentrations of the hemolyzed samples with nonhemolyzed samples, bias percentage was calculated by the formula:

$$[(CX - C1) / C1] \times 100.$$

C1: concentration of nonhemolyzed sample,

CX: concentration of hemolyzed sample.

All analyses were performed using Statistical Package for Social Sciences statistical package (SPSS,

version 15.0 for Windows XP). Wilcoxon signed-rank test analysis was used; P values < 0.05 were considered statistically significant.

Results

As a result of the mechanical trauma, the median free Hb concentrations for groups I, II, III, IV and V were measured as 0.02, 0.27, 0.75, 1.27 and 3.34 g/L, respectively. At free hemoglobin concentration 0.5 g/L, hemolysis was visible by the red color of the plasma. The results of the investigations were all presented in Table 1 and shown in Figure 1. Lactate dehydrogenase appeared to be most sen-

TABLE 1 Methods, median values of the analytes and free hemoglobin concentrations for each group, % bias of analyte concentrations in comparison to the nonhemolyzed group, and \pm acceptable limits of CLIA' 88, \pm desirable bias. Results with higher than CLIA limits and statistically significant differences are marked in bold.

Analyte	Method	No lysis	Free plasma hemoglobin, g/L				Desirable bias, \pm	CLIA, \pm
			0.27	0.75	1.27	3.34		
Albumin (g/L)	Bromcresol green	47.8	47.5 (-1.2%) P = 0.090	49.1 (+3.0%) P = 0.646	48.1 (+0.6%) P = 0.069	49.8 (+4.5%) P = 0.002	1.3%	10%
ALP (U/L, 37°C)	PNPP	88.0	90.5 (+0.6%) P = 0.473	83.5 (+8.5%) P = 0.034	96.0 (+9.1%) P = 0.592	71.5 (-9.5%) P = 0.050	6.4%	30%
ALT (U/L, 37°C)	UV without P5P	21	13.0 (+4.0%) P = 0.717	17.0 (+3.0%) P = 0.244	22.0 (+4.8%) P = 0.347	21.0 (+20%) P = 0.004	12%	20%
Amylase (U/L, 37°C)	CNP-triose	57	47.5 (-4.0%) P = 0.015	45.5 (-7.2%) P = 0.007	50 (-12.3%) P = 0.001	50 (-16.7%) P = 0.042	7.4%	30%
AST (U/L, 37°C)	UV without P5P	20.0	21.5 (+26.5%) P = 0.002	24.5 (+22.5%) P = 0.005	32 (+60%) P = 0.001	50 (+163.2%) P = 0.002	5.4%	20%
Bilirubin total (μ mol/L)	DPD, colorimetric	8.4	7.8 (-6.2%) P = 0.003	7.9 (-15.6%) P = 0.005	5.6 (-32.7%) P = 0.001	0.0 (-100%) P = 0.003	10.0%	20%
§ Calcium, total (mmol/L)	Arsenazo III	2.30	2.30 (+1.1%) P = 0.660	2.28 (-0.6%) P = 0.739	2.28 (-1.1%) P = 0.185	2.29 (0.0%) P = 0.348	0.8%	0.25 mmol/L
Chloride (mmol/L)	Ion-selective electrode	105	106.5 (-0.5%) P = 0.121	106.5 (-0.5%) P = 0.161	106 (+1.0%) P = 0.007	106.5 (+0.5%) P = 0.025	0.5%	5%
Cholesterol (mmol/L)	enzymatic, colorimetric	4.47	4.54 (-0.9%) P = 0.797	4.70 (+2.0%) P = 0.020	4.63 (+3.5%) P = 0.001	5.04 (+19.4%) P = 0.002	4.0%	10%

Analyte	Method	No lysis	Free plasma hemoglobin, g/L				Desirable bias, ±	CLIA, ±
			0.27	0.75	1.27	3.34		
CK (U/L, 37°C)	UV-NAC activated	80.5	81 (-1.2%) P = 0.057	86 (+6.8%) P = 0.005	97 (+20.1%) P = 0.002	112 (+16.7%) P = 0.005	11.5%	30%
Creatinine (µmol/L)	Jaffe mod/picrate	77.8	76 (-1.7%) P = 0.431	76.5 (0.0%) P = 0.783	77.8 (0.0%) P = 0.190	76.9 (-2.3%) P = 0.112	3.4%	15%
GGT (U/L, 37°C)	γ-Glutamyl-carboxy-nitroanilide	14	11 (-12%) P = 0.739	12.5 (-7.4%) P = 0.107	14 (0.0%) P = 0.942	17 (+25.9%) P = 0.014	10.8%	-
Glucose (mmol/L)	Hexokinase	4.77	4.94 (-1.1%) P = 0.032	4.36 (-4.9%) P = 0.012	4.72 (-1.2%) P = 0.001	4.58 (-8.3%) P = 0.003	2.2%	10%
HDL-C (mmol/L)	direct, non-immunologic	1.12	1.18 (+1.1%) P = 0.680	1.18 (+1.1%) P = 0.206	1.09 (-2.3%) P = 0.019	1.10 (-3.4%) P = 0.013	5.2%	30%
Iron (µmol/L)	Tri-piridil-triazin	14.3	14.2 (+1.9%) P = 0.188	17.5 (0.0%) P = 0.324	14.9 (+3.8%) P = 0.502	17.5 (+3.2%) P = 0.476	8.8%	20%
LDH (U/L, 37°C)	lactate-pyruvate, UV	279	356 (+21.1%) P = 0.002	475.5 (+66.0%) P = 0.005	684 (+145.2%) P = 0.001	1297 (+371.6%) P = 0.002	4.3%	20%
Magnesium (mmol/L)	Arsenazo I	0.82	0.82 (0.0%) P = 0.564	0.84 (+2.5%) P = 0.083	0.82 (0.0%) P = 0.001	0.90 (+12.8%) P = 0.002	1.8%	25%
Phosphorus (mmol/L)	Phospho-molybdate formation	0.90	0.90 (0.0%) P = 1.000	0.84 (+2.0%) P = 0.035	0.94 (+3.6%) P = 0.003	1.02 (+10.5%) P = 0.002	3.2%	10%*
§ Potassium (mmol/L)	ion-selective electrode	4.10	4.17 (+3.5%) P = 0.006	4.24 (+6.5%) P = 0.005	4.54 (+10.7%) P = 0.001	5.58 (+35.8%) P = 0.002	1.8%	0.5 mmol/L
Protein, total (g/L)	Biuret	74.7	73.1 (-1.4%) P = 0.504	72.5 (+0.4%) P = 0.183	75.2 (+0.7%) P = 0.944	77.1 (+3.1%) P = 0.002	1.2%	10%
§ Sodium (mmol/L)	ion-selective electrode	142	143 (+0.7%) P = 0.262	142.5 (-0.4%) P = 0.731	142 (0.0%) P = 0.377	140 (-1.4%) P = 0.205	0.3%	4 mmol/L
Triglycerides (mmol/L)	enzymatic, colorimetric	1.21	1.34 (-1.3%) P = 0.624	1.34 (-0.8%) P = 0.027	1.15 (-5.6%) P = 0.006	0.90 (-19.4%) P = 0.002	10.7%	25%
UIBC (µmol/L)	nitroso-PSAP	38.8	37.9 (+1.7%) P = 0.844	38.2 (+1.2%) P = 0.641	38.5 (-0.9%) P = 0.683	38.7 (+5.6%) P = 0.011	-	10%*
Urea (mmol/L)	urease, UV	4.51	4.43 (-1.9%) P = 0.206	4.26 (-1.9%) P = 0.763	4.51 (0.0%) P = 0.755	4.93 (+3.5%) P = 0.190	5.5%	9%
Uric acid (µmol/L)	uricase, UV	285.6	285.6 (-3.0%) P = 0.380	261.8 (-1.1%) P = 0.132	273.7 (-4.7%) P = 0.003	238 (-5.9%) P = 0.003	4.8%	17%

To compare the values, both CLIA'88 allowable limits and Analytical Quality Specifications for desirable bias (14,18) were given.

* according to the studies (2,8).

§ To compare the values of calcium with acceptable limits, % results were converted to mmol/L: 0.6% (0.05 mmol/L), 1.1% (0.025 mmol/L).

§ To compare the results of potassium with acceptable limits, % results were converted to mmol/L: 3.5% (0.14 mmol/L), 6.5% (0.26 mmol/L), 10.7% (0.44 mmol/L), 35.8% (1.47 mmol/L).

§ To compare the results of sodium with acceptable limits, % results were converted to mmol/L: 0.4% (0.50 mmol/L), 0.7% (1.00 mmol/L), 1.4% (2.00 mmol/L).

sitive to hemolysis; the increase of ~ 1000 U of lactate dehydrogenase per liter resulted in a 4.5-fold higher enzyme activity at 4.5 g of hemoglobin per liter of plasma than at 0.27 g/L. A considerable increase of 30 U/L in aspartate aminotransferase activity was a 2.5 fold increase in this enzyme's activity in severely hemolyzed plasma (hemoglobin < 4.5 g/L).

In electrolytes, potassium values were most affected, in direct proportion to the increased plasma free hemoglobin concentration, the average increase in potassium concentration being 1.48 mmol/L for hemoglobin concentrations up to 4.5 g/L, or about 1.4 fold (Figure 1).

Table 1 summarizes the influence of hemolysis on routine biochemistry testing. Differences are given as median values and percentage relative bias from the baseline specimens (no lysis). As expected, LD, AST and potassium values showed significant increases approximately linearly dependent on the free Hb concentrations in hemolyzed plasma. There was a decrease (100%) in total bilirubin concentration (Figure 1).

Values of inorganic phosphorus were significantly increased linearly with the increased free hemoglobin concentrations comparing with the baseline specimens. Clinically meaningful (approximately 1.2 fold) increases in enzyme activities were obser-

ved for GGT and ALT at 4.5 g of hemoglobin per liter of plasma.

Of the analytes evaluated, the bias values recorded for albumin, alkaline phosphatase, amylase, calcium, chloride, HDL-cholesterol, creatine kinase, creatinine, glucose, iron, magnesium, sodium, total protein, triglycerides, unsaturated iron binding capacity, urea and uric acid were lower than the CLIA allowable limits even in specimens containing up to 4.5 g/L of plasma hemoglobin although for some analytes variations were statistically significant ($P < 0.05$) (Table 1).

Although there was no given acceptable limit for GGT in CLIA' 88 considerations, we assumed the limit as 20% like most of the other enzymes. For the unsaturated iron binding capacity the evaluation was done according to some studies which was determined as $\pm 10\%$ (11,12).

Discussion

Analytic hemolysis interference occurs when the constituents of the plasma are at lower concentrations than the constituents in erythrocytes. The release of erythrocytic constituents can result in increased values for serum. Dilution is another possible effect especially for gross hemolysis, and may result in decreased values. While Hb absorbance peaks occur at ~417, 540, and 575 nm and at 415 nm (Soret wave) absorbs light very strongly, therefore at these wavelengths, spectrophotometric interference occurs due to an increase in the optical absorbance or a change in the blank value (8,11-13).

Cell contents at higher concentrations of potassium, phosphate, AST and LD enter the surrounding plasma when lysis of erythrocytes occurs (8,14). As expected in the current study AST, LD, and potassium values showed clinically meaningful increases like previous investigations (8,11-17,19). So we can say that even if macroscopically invisible hemolysis occurs, AST and LD measurements may contribute to high values.

Free hemoglobin with its pseudo-peroxidase activity interferes in the bilirubin procedure by inhibiting the diazonium color formation. In our study,




FIGURE 1. Interferogram for hemoglobin and measured parameters: LD, AST, potassium and total bilirubin.

we found statistically low values for all groups as in literature (11) and biochemical testing has to be done for this assay with a new sample at hemoglobin concentrations higher than 1 g/L.

For cholesterol, increase in concentration was significant at 2.5 g/L of Hb levels. According to a study, the interference may result from Hb decomposition of hydrogen peroxide. This effect can be compensated for by using an appropriate sample blank (15). For HDL-C, the interference was not clinically significant up to 10 g/L of hemoglobin concentration with the same assay method (20).

Although Hb is the major source of iron, hemolysis has a very little effect on the serum iron assay which may be caused by spectral overlap and by a chemical reaction between hemolysate and reaction components like ALP, GGT, amylase and iron assays (11,12,21).

We found a slight decrease in glucose and uric acid. This effect may become from a premature decomposition of hydrogen peroxide by Hb (15). Dilutional effect caused by the leakage of intracellular components into the surrounding fluid especially in severe hemolysis may cause lower values for glucose, sodium and calcium (14,19). In this study no lysate or any other solution was added in specimens which may also has a dilutional effect on the analysis.

Although CK is not a constituent of erythrocytes, intracellular adenylate kinase has been attributed to the interference in the CK assay (14,19). Correction can be done by adding inhibitors such as ade-

nosine monophosphate and diadenosine pentaphosphate, or by subtracting the activity measured in the absence of creatine phosphate (21).

Nowadays serum hemolysis index is a popular solution for interference detection preanalytically. Manufacturers give the list of test-specific serum indices for hemolysis, lipemia and bilirubin interferences. This can help laboratory staff to be aware of interference, study or reject the sample, and add comments to the results. But the standardization problem of the various analytical systems and different decision thresholds for various serum indices requires more effort (22).

Conclusion

We conclude that hemolysis affects plasma concentration of a whole range of biochemical parameters, whereas the most prominent effect of hemolysis is observed for AST, LD, potassium and total bilirubin. For other analytes; albumin, ALP, amylase, chloride, CK, HDL-cholesterol, glucose, magnesium, total protein, triglycerides, UIBC and uric acid, differences were statistically significant, but remained within the CLIA limits. We therefore recommend routine free Hb level determination in serum or plasma, or any other automated detection of the degree of hemolysis. Only for those analyses which are affected by the estimated degree of hemolysis, new samples have to be requested. Although time consuming, this procedure is highly useful, since it may help to avoid unnecessary rerun the samples and reduce the costs.

References

1. Bonini P, Plebani M, Ceriotti F, Rubboli F. Errors in laboratory medicine. *Clin Chem* 2002;48:691-8.
2. Plebani M, Carraro P. Mistakes in a stat laboratory: types and frequency. *Clin Chem* 1997;43:1348-51.
3. Carraro P, Servidio P, Plebani M. Haemolyzed specimens: a reason for rejection or clinical challenge? *Clin Chem* 2000;46:306-7.
4. Hinckley C. Defining the best quality-control systems by design and inspection. *Clin Chem* 1997;43:873-9.
5. Glick M, Ryder K, Jackson S. Graphical comparisons of interferences in clinical chemistry instrumentation. *Clin Chem* 1986;32:470-5.
6. Plebani M. Laboratory errors: How to improve pre-and post-analytical phases? *Biochem Med* 2007;17:5-9.
7. Simundic AM, Nikolac N, Ivankovic V, Ferenec-Ruzic D, Magdic B, Kvaternik M, Topic E. Comparison of visual vs.automated detection of lipemic, icteric and hemolyzed specimens: can we rely on a human eye? *Clin Chem Lab Med* 2009;47:1361-5.

8. Thomas L. Haemolysis as influence and interference factor. *eJIFCC* vol 13 no 4: Available at: <http://www.ifcc.org/ejifcc/vol13no4/130401002.htm>. Accessed: September 25, 2010.
9. Fairbanks VF, Klee GG. Biochemical aspects of hematology. In: Burtis CA, Ashwood ER, eds. *Tietz Textbook of Clinical Chemistry*. 3rd ed. Philadelphia: WB Saunders Company; 1999. p.1673-6.
10. *Clinical Laboratory Improvements Amendments of 1988. Final Rule. Laboratory Requirements. Federal Register* 1992;57:7002-288.
11. Sonntag O. Haemolysis as an interference factor in clinical chemistry. *J Clin Chem Clin Biochem* 1986;24:127-39.
12. Jay DW, Provasek D. Characterization and mathematical correction of hemolysis interference in selected Hitachi 717 assays. *Clin Chem* 1993;39:1804-10.
13. Grafmeyer D, Bondon M, Manchon M, Levillain P. The influence of bilirubin, haemolysis and turbidity on 20 analytical tests performed on automatic analysers. Results of an interlaboratory study. *Eur J Clin Chem Clin Biochem* 1995;33:31-52.
14. Lippi G, Salvagno LG, Montagnana M, Brocco G, Guidi GC. Influence of hemolysis on routine clinical chemistry testing. *Clin Chem Lab Med* 2006;44:311-6.
15. Yucel D, Dalva C. Effect of in vitro hemolysis on 25 common biochemical tests. *Clin Chem* 1992;38:575-7.
16. Scott MG, Heusel JW, Le Grys VA, Siggaard AO. In: Burtis CA, Ashwood ER, eds. *Tietz Textbook of Clinical Chemistry*. 4th ed. Philadelphia: WB Saunders Company; 2006. p.1058-60.
17. Randall AG, Garcia-Webb P, Beilby JB. Interference by haemolysis, icterus and lipaemia in assays on the Beckman Synchron CX5 and methods for correction. *Ann Clin Biochem* 1990;27:345-52.
18. Ricos C, Alvarez V, Cava F, Garcia-Lario JV, Hernandez A, Jimenez CV, et al. Current databases on biologic variation: pros, cons and progress. *Scan J Clin Lab Invest* 1999;59:491-500.
19. Frank JJ, Bermes EW, Bickel MJ, Watkins BF. Effect of in vitro hemolysis on chemical values for serum. *Clin Chem* 1978;24: 1966-70.
20. Nauck M, März W, Wieland H. New immunoseparation-based homogeneous assay for HDL-cholesterol compared with three homogeneous and two heterogeneous methods for HDL-cholesterol. *Clin Chem* 1998;44:1443-51.
21. Szasz G, Gerhardt W, Gruber W, Bernt E. Creatine kinase in serum: 2. Interference of adenylate kinase with the assay. *Clin Chem* 1976;2:1806-11.
22. Simundic AM, Topic E, Nikolac N, Lippi G. Hemolysis detection and management of hemolysed specimens. *Biochemia Medica* 2010;20:154-9.

Učinci hemolize na rutinsko određivanje biokemijskih parametara

Sažetak

Uvod: Hemoliza je još uvijek najčešći razlog odbijanja uzoraka, iako je ponovno dobivanje uzorka velik problem. Cilj ovog istraživanja je ispitati učinke hemolize u njenim raznim stupnjevima za najčešće određivane biokemijske parametre kako bi se spriječila nepotrebna odbijanja.

Metode i ispitanici: U ispitivanje je uključeno šesnaest zdravih dobrovoljaca. Prema koncentracijama hemoglobina određena su četiri stupnja hemolize, a ispitanici su podijeljeni u pet skupina: Skupina I: 0-0,10 g/L, Skupina II: 0,10-0,50 g/L, Skupina III: 0,51-1,00 g/L, Skupina IV: 1,01-2,50 g/L, Skupina V: 2,51-4,50 g/L. Liza eritrocita je postignuta mehaničkom traumom.

Rezultati: Interferencija uslijed hemolize utjecala je na određivanje aktivnosti laktat-dehidrogenaze (LD) i aspartat-aminotransferaze (AST) kod gotovo neprepoznate hemolize pri vizualnom pregledu (koncentracija hemoglobina u plazmi < 0.5 g/L). Primijećene su klinički značajne varijacije koncentracije kalija i ukupnog bilirubina u umjereno hemoliziranim uzorcima (hemoglobin > 1 g/L). Kod određivanja aktivnosti alanin-aminotransferaze (ALT), gama-glutamilttransferaze (GGT) te koncentracije kolesterola i anorganskog fosfata (P) nisu primijećene interferencije sve do stupnjeva jake hemolize (hemoglobin: 2,5-4,5 g/L). Razlike u aktivnostima alkalne fosfataze (ALP), amilaze, kreatinin kinaze (CK) te koncentracijama albumina, klorida, HDL-kolesterola, glukoze, magnezija, ukupnog proteina, triglicerida i mokraćne kiseline te u nezasićenom kapacitetu vezivanja željeza (UIBC) bile su statistički značajne, no unutar dozvoljenih granica prema CLIA kriterijima (engl. *Clinical Laboratory Improvement Amendments*, CLIA).

Zaključak: Kako bi se izbjegla prijeanalitička metoda vizualnog pregleda uzorka za prepoznavanje hemolize, neodgovarajuće odbijanje uzoraka i/ili ponovno testiranje uslijed hemolize preporuča se rutinsko određivanje koncentracije slobodnog hemoglobina u plazmi ili serumu, kao što pokazuje ovo istraživanje. Za određivanje koncentracije ili aktivnosti analita koji često podliježu interferenciji zbog hemolize potrebno je zatražiti novi uzorak.

Ključne riječi: hemoliza; pogriješke u prijeanalitičkoj fazi; interferencija; analiti